



Sample Information

Date and Time : 2017-04-21 13:42:14  
User : CHAO  
Sample : 488707  
Inj. Volume : 1  
MW : 818.0

Probe: ESI  
Nebulizer Gas Flow: 1.5L/min  
CDL: -20.0v  
CDL Temp: 250°C  
Block Temp: 200°C

Probe bias: -3.5kv  
Detector: 1.0kv  
T. Flow: 0.2ml/min  
B. conc: 50%H2O/50%ACN